syngo® Dynamics

DICOM Conformance Statement

VA40

siemens-healthineers.com

SIEMENS Healthineers

This page is left blank intentionally.

Table of Contents

1 Introduction 5

Purpose of this Document 5
Limitations 7
Sources for this Document 8
Acronyms and Abbreviations 8

2 Media Storage Conformance Statement 9

Introduction 9
Implementation Model 9
Application Entity Specifications 13
Private Elements Exported to Media 21
Augmented and Private Application Profiles 21
Extensions/Specializations/Privatizations 22
Configuration 22
Support of Extended Character Sets 22

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

Table of Contents

This page is left blank intentionally.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

1 Introduction

1.1 Purpose of this Document

This document is a DICOM Conformance Statement for the *syngo*® Dynamics Workplace.

The *syngo*® Dynamics Workplace provides diagnostic display of DICOM Images conforming to the Ultrasound Image Storage, Ultrasound Multi-frame Image Storage, X-ray Angiographic Image Storage, Nuclear Medicine Image Storage, Positron Emission Tomography Image Storage and Secondary Capture Image Storage SOP Classes. Images are reviewed through the DICOM Image Review component which displays DI-COM images without any additional compression. The *syngo*® Dynamics Workplace is part of the Siemens syngo® Suite and is designed to operate in conjunction with the *syngo*® Dynamics Server. The *syngo*® Dynamics Workplace uses DICOM as an external interface standard. Non-DICOM internal communication mechanisms between the *syngo*® Dynamics Workplace and the *syngo*® Dynamics Server are used for the transfer and display of images.

The *syngo*® Dynamics Workplace provides support for the DICOM US Region Calibration Module. When ultrasound images contain this module, the *syngo*® Dynamics Workplace is able to interpret the region calibration data provided by the Ultrasound acquisition unit. Standard pixel spacing attributes are also supported if provided with Ultrasound captures. The *syngo*® Dynamics Workplace provides support for the DI-COM Imager Pixel Spacing and Estimated Radiographic Magnification Factor attributes often sent with X-Ray Angiographic captures. When XA captures contain this data the *syngo*® Dynamics Workplace can interpret them. In the case where no calibration data is provided by the acquisition device, captures may be manually calibrated using the *syngo*® Dynamics Workplace.

1.1.1 Image Display Support (Software)

The *syngo*® Dynamics Workplace displays Ultrasound, NM, PT, SC and XA Angiographic images from the *syngo*® Dynamics Server or DICOM Media.

The *syngo*® Dynamics Workplace supports the image formats shown in “Table 1. 1 Software Review Supported Image Formats” when displaying images using software. The *syngo*® Dynamics Workplace supports the display of 8, 10, 12 and 16 bit image formats.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

1 Introduction

Photometric Interpretation	Transfer Syntax	Planar Configuration
MONOCHROME1	Uncompressed Implicit VR Little Endian 1.2.840.10008.1.2	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	N/A
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.57	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	N/A
MONOCHROME2	Uncompressed Implicit VR Little Endian 1.2.840.10008.1.2	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	N/A
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.57	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	N/A

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

Photometric Interpretation	Transfer Syntax	Planar Configuration
RGB	Uncompressed Implicit VR Little Endian 1.2.840.10008.1.2	0 - Chunky
	Uncompressed Implicit VR Little Endian 1.2.840.10008.1.2	1 - Planar
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	0 - Chunky
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	1 - Planar
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	0 - Chunky
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	1 - Planar
YBR_FULL	Uncompressed Implicit VR Little Endian 1.2.840.10008.1.2	0 - Chunky
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	0 - Chunky
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	0 - Chunky
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	1 - Planar
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	0 - Chunky
YBR_FULL_422	Uncompressed Implicit VR Little Endian 1.2.840.10008.1.2	0 - Chunky
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	0 - Chunky
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	0

Table 1. 1 Software Review Supported Image Formats

1.2 Limitations

None

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

1 Introduction

1.3 Sources for this Document

ACR-NEMA Digital Imaging and Communications in Medicine (DICOM) Version 3 Parts 1 - 20.

1.4 Acronyms and Abbreviations

The following Acronyms and abbreviations are used in this document:

- ACR - American College of Radiology
- AE - Application Entity
- DICOM - Digital Imaging and Communications in Medicine
- FSC - File Set Creator
- FSR - File Set Reader
- FSU - File Set Updater
- NEMA - National Electrical Manufacturers Association
- SCP - Service Class Provider
- SCU - Service Class User
- SOP - Service Object Pair
- TCP/IP - Transmission Control Protocol/Internet Protocol

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement

2.1 Introduction

This section specifies the *syngo*® Dynamics Workplace compliance to the DICOM Media Interchange. It details the DICOM Media Storage Application Profiles and roles which are supported.

This station provides DICOM interchange capabilities on CD, DVD as well as regular Windows file systems and USB media with different application profiles supported for each media type. Support for CD/DVD media is dependent on the appropriate hardware being installed on the station.

Note that the format of this section follows the format described in DICOM Standard PS 3.2 (Conformance). Please refer to that part of the standard while reading this section.

2.2 Implementation Model

2.2.1 Application Data Flow Diagram

The Basic and Specific Application models for writeable and non-writeable media are shown in the following illustrations:

2.2.1.1 Description of the Data Flow Diagram for Writable Media.

The Display/Edit Application Entity (AE) handles the Directory Display, Image Viewing, Study Updating, Study Copying and Media Creation functionality for the writable media device. The Display/Edit Application Entity (AE) is commanded by the user to perform DICOM Services operating on the DICOM media through the use of buttons and menu selections on the graphical user interface of the Workplace.

The Application models for writeable media devices are shown in the figures below.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement



Figure 2.1 CD-R Implementation Model

The Application models for the CD-ROM/DVD-ROM device are shown in the below figure.



Figure 2.2 CD-ROM and DVD-ROM Implementation Model

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2.2.1.2 Description of the Data Flow Diagram for the CD-RW/DVD-RW Device.

The Display/Edit Application Entity (AE) handles the Directory Display, Image Viewing, Study Updating, Study Copying and CD/DVD Creation functionality for the CD/DVD device. The Display/Edit Application Entity (AE) is commanded by the user to perform DICOM Services operating on the DICOM media through the use of buttons and menu selections on the graphical user interface of the station. The Application models for the CD/DVD device are shown in Figure 2.4. Note that the *syngo*® Dynamics Workplace does not support any official Application Profiles for CD/DVD. However, DICOM Media exchange files can be written and read along with a DI-COMDIR in accordance with Figure 2.3.



Figure 2.3 CD-RW and DVD-RW Implementation Model

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement



Figure 2.4 USB - Windows File System Implementation Model

2.2.2 Functional Definition of Application Entities

The *syngo*® Dynamics Workplace has only one Application Entity: the Display/Edit Application. The Display/Edit Application supports the following functions:

- Display a directory listing of the DICOM File Set (FSR)
- Display images from a DICOM File Set (FSR)
- Copy images from a DICOM File Set (FSR)
- Update or Delete DICOM File Sets (FSU)
- Create DICOM File Set on a CDROM/DVD (FSC)
- Create DICOM File Set on USB media or Windows file system location (FSC)

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2.2.3 Sequencing of Real World Activities

Users of the *syngo*® Dynamics Workplace initiate actions which trigger the reading, writing and updating of DICOM objects to and from the supported media.

A DICOM File Set must exist on the media for a DICOM File Set to be updated.

2.2.4 File Meta Information for Implementation Class and Version

The *syngo*® Dynamics Workplace Display/Edit Application uses the following implementation identifying parameters:

- File Meta Information Version 1
- Implementation Class UID 1.3.12.2.1107.5.8.11.105

2.3 Application Entity Specifications

2.3.1 Display/Edit Application Entity Specification

The Display/Edit Application Entity provides standard conformance to the DICOM Interchange Option of the Media Storage Service Class. The Application Profiles and roles are listed in the below table

Supported AP's	Real-World Activity	Roles	SC Option
STD-US-ID-MF- CDR	Display Directory	FSR	Interchange
STD-US-SC-MF-CDR			
STD-XABC-CD			
STD-XA1K-CD STD-XA1K-DVD			
STD-CTMR-CD			
STD-CTMR-DVD			
STD-GEN-DVD-JPEG			
STD-GEN-USB-JPEG			
STD-GEN-USB-JPEG-2000			

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement

Supported AP's	Real-World Activity	Roles	SC Option
STD-XABC-CD STD-XA1K-CD STD-XA1K-DVD STD-CTMR-CD STD-CTMR-DVD STD-GEN-DVD-JPEG STD-GEN-USB-JPEG STD-GEN-USB-JPEG-2000	View Images[a]	FSR	Interchange
STD-XABC-CD STD-XA1K-CD STD-XA1K-DVD STD-CTMR-CD STD-CTMR-DVD STD-GEN-DVD-JPEG STD-GEN-USB-JPEG STD-GEN-USB-JPEG-2000 STD-US-ID-MF-CDR STD-US-SC-MF-CDR STD-XABC-CD STD-XA1K-CD STD-XA1K-DVD STD-CTMR-CD STD-CTMR-DVD STD-GEN-DVD-JPEG	Copy to Local Storage	FSR	Interchange
CDR STD-US-SC-MF-CDR STD-XABC-CD STD-XA1K-CD STD-XA1K-DVD STD-CTMR-CD STD-CTMR-DVD STD-GEN-DVD-JPEG	Update Studies	FSU	Interchange
STD-US-SC-MF-CDR STD-XABC-CD STD-XA1K-CD	Create CD-R	FSC	Interchange
STD-XA1K-DVD STD-CTMR-DVD STD-GEN-DVD-JPEG	Create DVD	FSC	Interchange

a. Partial Conformance - See "Real World Activity: View Images"

Table 2. 1 Display/Edit Application Entity Profiles, Real-World Activities, and Roles

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

Media Storage Conformance Statement 2

The Display/Edit Application Entity provides support for Media Storage SOP Classes shown in the below table.

Information Object Definition	SOP Class UID	Transfer Syntax	Transfer Syntax UID
DICOM Media Storage Directory	1.2.840.10008.1.3.10	Explicit VR Little Endian	
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	Explicit VR Little Endian1	1.2.840.10008.1.2.
		JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.70
Multi-frame True Color Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.4	Explicit VR Little Endian	1.2.840.10008.1.2.1
Digital X-Ray Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.1.1	JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.70
		JPEG Lossless	1.2.840.10008.1.2.4.70
Digital X-Ray Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Explicit VR Little Endian	1.2.840.10008.1.2.1
		JPEG Lossless	1.2.840.10008.1.2.4.70
CT Image	1.2.840.10008.5.1.4.1.1.2	Explicit VR Little Endian	1.2.840.10008.1.2.1
		JPEG Lossless	1.2.840.10008.1.2.4.70
Ultrasound Multi- frame Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Explicit VR Little Endian	1.2.840.10008.1.2.1
MR Image	1.2.840.10008.5.1.4.1.1.4	Explicit VR Little Endian	1.2.840.10008.1.2.1
MR Image Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.4 1.2.840.10008.5.1.4.1.1.6.1	JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.70
		Explicit VR Little Endian1	1.2.840.10008.1.2.1
Ultrasound Image Storage SC Image (grayscale)	1.2.840.10008.5.1.4.1.1.6.1 1.2.840.10008.5.1.4.1.1.7	RLE Lossless Image Compression	1.2.840.10008.1.2.5
		JPEG Lossy, Baseline	1.2.840.10008.1.2.4.50
		Explicit VR Little Endian	1.2.840.10008.1.2.1
SC Image (grayscale) SC Image (palette color)	1.2.840.10008.5.1.4.1.1.7 1.2.840.10008.5.1.4.1.1.7	JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.70
		Explicit VR Little Endian1	1.2.840.10008.1.2.1

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement

Information Object Definition	SOP Class UID	Transfer Syntax	Transfer Syntax UID
SC Image (palette color) X-Ray Angiographic ge	1.2.840.10008.5.1.4.1.1.7 1.2.840.10008.5.1.4.1.1.12.1	JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.70
		JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.57
X-Ray Angiographic Image	1.2.840.10008.5.1.4.1.1.12.1	JPEG Lossless (Process 14)	1.2.840.10008.1.2.4.70

Table 2.2 Supported Media Storage SOP Classes

2.3.1.1 File Meta Information for Display/Edit Application Entity

The *syngo*® Dynamics Workplace Source Application Entity Title will be the AE Title assigned to the *syngo*® Dynamics Workplace.

2.3.1.2 Real World Activities for the Display/Edit Application Entity

The *syngo*® Dynamics Workplace Display/Edit Application Entity is used for the following real world activities:

- Display Directory Listing - In this activity the Display/Edit Application Entity acts as a File-Set Reader.
- Viewing of Images - In this activity the Display/Edit Application Entity acts as a File-Set Reader.
- Copy to Local Storage - In this activity the Display/Edit Application Entity acts as a File- Set Reader.
- Updating Images - In this activity the Display/Edit Application Entity acts as a File-Set Updater.
- Creating a DVD, or CD-R - In this activity the Display/Edit Application Entity acts as a File-Set Creator.
- Creating a Windows file system or USB media set - In this activity the Display/Edit Application Entity acts as a File-Set Creator.

2.3.1.2.1 Real World Activity: Display Directory

The *syngo*® Dynamics Workplace Display/Edit Application is an FSR when reading the directory of the medium. This will result in an overview of the patients, studies and images in the *syngo*® Dynamics Workplace Study List.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

Media Storage Application Profile for the RWA: View Images

For the list of Application Profiles that invoke this AE for the View Images RWA, see "Table 2. 1 Display/Edit Application Entity Profiles, Real-World Activities, and Roles". There are no extensions or specializations.

2.3.1.2.2 Real World Activity: View Images

The *syngo*® Dynamics Workplace Display/Edit Application is an FSR when viewing images from the medium.

The *syngo*® Dynamics Workplace partially supports the multi-frame ultrasound "image display" Application Profile for MOD and the multi-frame ultrasound "spatial calibration" Application Profile for DVD, CD-R, Windows File System and USB media. The below table shows the supported image formats when viewing studies from re-movable media or fixed Windows file system locations.

Photometric Interpretation	Transfer Syntax	Bits Allocated	Bits Stored	Samples Per Pixel	Planar Configuration
MONOCHROME1	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	1	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	16	16	1	N/A
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.57	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	8	8	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	1	N/A

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement

MONOCHROME2	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	1	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	16	16	1	N/A
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.57	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	8	8	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	1	N/A
RGB	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	3	0 - Chunky
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	3	1 - Planar
PALETTE COLOR	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	1	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	16	16	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	16	16	1	N/A
YBR_FULL	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	3	0 - Chunky
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	3	0 - Chunky
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	3	0 - Planar
YBR_FULL_422	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	8	8	3	0 - Planar

Table 2.3 Media Review Supported Image Formats

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

Media Storage Application Profile for the RWA: View Images

For the list of Application Profiles that invoke this AE for the View Images RWA, see "Table 2. 1 Display/Edit Application Entity Profiles, Real-World Activities, and Roles". There are no extensions or specializations.

2.3.1.2.3 Real World Activity: Update Studies

The *syngo*® Dynamics Workplace Display/Edit Application is an FSU using the Interchange option when adding studies to the medium. The Dis-play/Edit Application will copy SOP Instances from Local Storage to the medium.

The Display/Edit Application cannot delete studies from media.

The Display/Edit Application cannot add studies to media that has been write-protected.

Media Storage Application Profile for the RWA: Update Studies

For the list of Application Profiles that invoke this AE for the Copy to Local Storage RWA, see "Table 2. 1 Display/Edit Application Entity Profiles, Real-World Activities, and Roles". There are no extensions or specializations.

2.3.1.2.4 Real World Activity: Create CD-R/DVD

The *syngo*® Dynamics Workplace Display/Edit Application is an FSC when creating a CD- R/DVD. A DICOMDIR is created and studies can be exported to the CD-R/DVD (See RWA: Update Studies).

Media Storage Application Profile for the RWA: Create CD-R/DVD

For the list of Application Profiles that invoke this AE for the Create CD-R RWA, see "Table 2. 1 Display/Edit Application Entity Profiles, Real-World Activities, and Roles". There are no extensions or specializations.

2.3.1.2.5 Real World Activity: Display Directory

The *syngo*® Dynamics Workplace Display/Edit Application is an FSR when reading the directory of the medium. This will result in an overview of the patients, studies and images in the *syngo*® Dynamics Workplace Study List.

Media Storage Application Profile for the RWA: Display Directory

For the list of Application Profiles that invoke this AE for the Display Directory RWA, see "Table 2. 1 Display/Edit Application Entity Profiles, Real-World Activities, and Roles". There are no extensions or specializations.

2.3.1.2.6 Real World Activity: View Images

The *syngo*® Dynamics Workplace Display/Edit Application is an FSR when viewing images from the medium.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2 Media Storage Conformance Statement

The *syngo*® Dynamics Workplace partially supports the multi-frame ultrasound "image display" Application Profile for MOD and the multi-frame ultrasound "spatial calibration" Application Profile for DVD, and CD-R. The below table shows the supported image formats when viewing studies from removable media.

Photometric Interpretaton	Transfer Syntax	Bits Allocated	Bits Stored	Samples Per Pixel	Planar Configuration
MONOCHROME1	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	1	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	16	16	1	N/A
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.57	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	8	8	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	1	N/A
MONOCHROME2	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	1	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	16	16	1	N/A
	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.57	8	8	1	N/A
	JPEG Lossless, Non-Hierarchical (Process 14) 1.2.840.10008.1.2.4.70	8	8	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	1	N/A

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

RGB	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	3	0 - Chunky
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	3	1 - Planar
PALETTE COLOR	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	1	N/A
	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	16	16	1	N/A
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	16	16	1	N/A
YBR_FULL	Uncompressed Explicit VR Little Endian 1.2.840.10008.1.2.1	8	8	3	0 - Chunky
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	3	0 - Chunky
	RLE Lossless Image Compression 1.2.840.10008.1.2.5	8	8	3	0 - Planar
YBR_FULL_422	JPEG Lossy Baseline (Process 1) 1.2.840.10008.1.2.4.50	8	8	3	0 - Planar

Table 2.4 Media Review Supported Image Formats

2.4 Private Elements Exported to Media

The *syngo*® Dynamics Workplace may not export all private elements of certain US SOP class objects if the private elements contain raw image data.

2.5 Augmented and Private Application Profiles

The *syngo*® Dynamics Workplace has no augmented or private Application Profiles.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2.6 Extensions/Specializations/Privatizations

The *syngo*® Dynamics Workplace has no extensions, specializations or privatizations of SOP Classes and Transfer Syntaxes.

2.7 Configuration

The Source AE Title encoded in the File Meta Information is derived from the AE Title of the workstation hosting the Workplace application.

2.8 Support of Extended Character Sets

The *syngo*® Dynamics Workplace supports the following character sets:

- ISO-IR 6 (default) - Default repertoire
- ISO-IR 100 - Latin Alphabet No. 1

The *syngo*® Dynamics Workplace does not support multi-byte characters.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

This page is left blank intentionally.

© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.

2019

Manufacturer's note:

All product designations and company names are trademarks or registered trademarks of the corresponding companies.

Siemens Healthineers reserves the right to modify the design and specifications contained herein without prior notice.

Some of the specifications described herein may not be currently available in all countries. Please contact your local Siemens Healthineers Sales representative for the most current information.

Caution: US federal law restricts the herein described devices to sale by or on the order of a physician.

The software described herein is CE compliant in accordance with Directive MDD 93/42/EEC Annex II.

The original language of this document is English.

Origin US

Local Contact Information in the USA
Siemens Medical Solutions USA, Inc.
40 Liberty Boulevard
Malvern, PA, 19355-9998
U.S.A
Telephone:
+1-610-448-4500
1-888-826-9702

Siemens Healthineers Headquarters
Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany
Phone: +49 9131 84-0
siemens-healthineers.com

Legal Manufacturer
Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany

Published by Siemens Healthcare GmbH
© Siemens Healthcare GmbH / Siemens Medical Solutions USA, Inc. 2019. All rights reserved.
Date of first issue: 2019-11